

DBH
mmB

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/26/2014**

Instrument Serial Number: 010557

As of 26-Jun-14

Start Date	End Date	Agency Name
24-Jun-14		DFS Central Lab
07-Jan-09	24-Jun-14	Richmond County Sheriff's Office
15-Aug-08	07-Jan-09	DFS Central Lab

DBH
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010557 Worksheet Start Date 6/24/2014
Location Northern Neck Regional Jail
Address 3908 Richmond Rd Warsaw VA 22572
DFS Technician Donald Hall License No. 21250

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 760 Reference Barometer (mm HG) 762
Reference Barometer(RB)Serial # 009109 RB Calibration Due 8/1/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.302
0.300	0.300	0.291	0.309	Sample 2	0.302
Precision		sample min	sample max	Sample 3	0.301
0.001		0.301	0.302		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.101
0.100	0.100	0.097	0.103	Sample 2	0.101
Precision		sample min	sample max	Sample 3	0.101
0		0.101	0.101		

Dry gas standard Lot No. (with tank no.) AG227501-22

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010557

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B. Hall

Date

6/26/14

Issuing Analyst

Michelle Buller

Date

6/26/14msk

YBH
mmB




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 06/24/2014
TEST EQUIPMENT NUMBER 010557		

RESULTS: TIME SAMPLE TAKEN 12:58 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 25 Jun 2014 at 16:49

DBH
mmB

Test Results

Instrument Serial Number 010557

Test # 000332 Subject Test

Test Location 1 Department of
Test Date 24 Jun 2014

Test Location 2 Forensic Science
Test Time 12:51
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

Effective Date 10/01/2012

License Number 21250

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 24 Jun 2014 End Time 12:59

Result Time 12:58

Result Date 24 Jun 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:51

Data Type BLK

Sample Value 0.000

Sample Time 12:52

Data Type CHK

Sample Value 0.101

Sample Time 12:53

Data Type BLK

Sample Value 0.000

Sample Time 12:54

Data Type SUBJ

Sample Value 0.000

Sample Time 12:55

Data Type BLK

Sample Value 0.000

Sample Time 12:56

Data Type SUBJ

Sample Value 0.000

Sample Time 12:58

Data Type BLK

Sample Value 0.000

Sample Time 12:59

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG227501-22

Standard Expiration Date 10/01/2014

Tank Pressure 340

Barometric Pressure 759 mmHg

Blow Sample Number 1 Blow Duration 3.78 sec

Blow Volume 1895 cc End-of-Blow Time 12:55

Blow Sample Number 2 Blow Duration 3.80 sec

Blow Volume 1933 cc End-of-Blow Time 12:58

Tamper Evident Stamp 2cca2036

Test Status Code 0

Test Status Success

Intox EC/IR-II: Accuracy Check

D64
mmB

Department of Forensic Science

Serial Number: 010557 Test Number: 333

Test Date: 06/24/2014 Test Time: 13:01 EDT

Dry Gas Target: 0.100

Lot Number: AG227501-22 Exp Date: 10/01/2014

Tank Pressure: 340 psi Barometric Pressure: 759 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:01
CHK	0.101	13:02
BLK	0.000	13:04
CHK	0.101	13:05
BLK	0.000	13:06
CHK	0.101	13:07

Test Status: *Success*

Calibration CRC: 3E7EEA90

Intox EC/IR-II: Accuracy Check

DB/H
mmB

Department of Forensic Science

Serial Number: 010557 Test Number: 334

Test Date: 06/24/2014 Test Time: 13:09 EDT

Dry Gas Target: 0.300

Lot Number: AG334603-07 Exp Date: 12/12/2015

Tank Pressure: 203 psi Barometric Pressure: 759 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:10
CHK	0.302	13:11
BLK	0.000	13:12
CHK	0.302	13:13
BLK	0.000	13:15
CHK	0.301	13:16

Test Status: *Success*

Calibration CRC: 3E7EEA90